

From Eye to Insight



OPERATING MANUAL EM GP2

Version 11/17



Important Note

Leica reserves the right to change technical specifications as well as manufacturing processes without prior notice. Only in this way is it possible to continuously improve the technology and manufacturing techniques used to provide our customers with excellent products.

Any copyrights of this document are retained by Leica Mikrosysteme GmbH, Vienna. Any reproduction of text and illustrations (or any parts thereof) by printing, photocopying, or other methods (including electronic systems and media) requires express prior permission in writing.

Issued by:

Leica Mikrosysteme GmbH
Hernalser Hauptstrasse 219
A-1170 Vienna

EM GP2

Operating Manual

Leica EM GP2 Serial Number:

Date of purchase:

For the instrument serial number, please refer to the name type label on the back of the instrument!



**Please read this instruction manual carefully
before operating the instrument.
For Research use only!**

IMPORTANT

READ CAREFULLY BEFORE USE

KEEP FOR FUTURE REFERENCE

Foreword

This manual is intended to provide essential information about the proper operation and servicing of the product. Users must familiarize themselves with the system and its components before commissioning. Particular attention must be paid to the safety instructions. This product may not be used beyond the limits specified in the technical data sheet.

This manual is for informational purposes and cannot cover all possible configurations and processes performed by the product, nor present guidelines to counter every possible contingency during its operation and maintenance. Users are encouraged to contact service or their sales representative when having additional questions.

Warranty claims will only be considered if the product is used according to the guidelines specified in this operating manual. See Section 8 for further warranty information. In addition, all generally valid, legal, and otherwise binding regulations for accident and environmental protection must be observed and communicated.

Leica reserves the right to change technical specifications and manufacturing processes without prior notice. Only in this way, Leica can continuously improve the technology and techniques used to provide its customers with excellent products.

Any copyrights of this document are retained by Leica Mikrosysteme GmbH, Vienna. Any reproduction of text and illustrations (or any parts thereof) by printing, photocopying, or other methods (including electronic systems and media) as well as handing over to third parties requires approval of Leica Mikrosysteme GmbH.

The original operating manual was written in English. In case of discrepancies, versions of this manual in other languages must refer to the English version.




Table of Contents

1	Introduction.....	1
1.1	Product Identification	1
1.2	Field of Application and Intended Use	1
1.3	Qualification of Operating Personnel.....	2
1.4	Symbols and Warnings.....	2
1.5	General Safety Information.....	3
1.6	Safety Measures regarding Nitrogen	4
1.6.1	Physical Properties	4
1.6.2	Safe Storage	5
1.6.3	Equipment and Handling.....	6
1.7	Safety Measures regarding Ethane and Propane	8
1.7.1	Physical Properties	8
1.7.2	Safe Storage	8
1.7.3	Equipment and Handling.....	8
2	Product Description	10
2.1	General Overview.....	10
2.2	Dimensions and Weight.....	12
2.2.1	Product Dimensions	12
2.2.2	Stereomicroscope	13
2.3	Consumption and Emissions	14
2.4	Protective Measures	14
2.4.1	Constructive Protective Measures	14
2.4.2	Technical Protective Measures.....	14
2.4.3	Informative Protective Measures	15
2.5	Type Label.....	16
3	Transportation and Storage.....	17
3.1	Transportation	17
3.2	Relocation.....	17
3.3	Storage	17
4	Installation	18
4.1	Work Area Conditions.....	18
4.2	Unpacking.....	19
4.3	Assembly	20
4.4	Accessories	21

5	Operation.....	22
5.1	Switching on and User Interface.....	22
5.2	Parameters.....	24
5.2.1	Environment Parameters.....	24
5.2.2	Load Specimen Parameters.....	27
5.2.3	Blot Parameters.....	29
5.2.4	Plunge / Transfer Parameters.....	34
5.3	Programs.....	36
5.3.1	Loading Programs.....	36
5.3.2	Editing Programs.....	37
5.3.3	Copying Programs.....	37
5.3.4	Deleting Programs.....	38
5.4	Preparing the Instrument.....	38
5.4.1	Filter Paper.....	38
5.4.2	Humidifier.....	40
5.4.3	Emptying the Drip Tray.....	42
5.4.4	Freezing Chamber and Cool-Down.....	43
5.4.5	Filling the Secondary Cryogen.....	45
5.5	Freezing a Grid.....	46
5.5.1	Attaching the Forceps/Grid.....	46
5.5.2	Lower Chamber.....	48
5.5.3	Adjust Parameters.....	49
5.5.4	Applying the Specimen.....	49
5.5.5	Blotting.....	50
5.5.6	Freezing.....	51
5.5.7	Transfer to Cryo Grid Box.....	51
5.6	Transferring a Cryo-Grid Box.....	52
5.7	Shutting Down and Bake Out.....	53
5.8	Settings.....	56
5.8.1	Tab “Process”.....	56
5.8.2	Tab “System”.....	57
6	Maintenance and Repairs.....	58
6.1	Service Menu.....	58
6.2	Software Updates.....	58
6.2.1	Preparing for the Update.....	58

6.2.2	Performing the Update	59
6.3	Cleaning and Upkeep	61
6.3.1	Cleaning the Humidifier Tank	61
7	Disposal	63
7.1	Decommissioning	63
7.2	Disposal	63
8	Troubleshooting	64
8.1	About this Manual	64
8.2	Warranty	64
8.3	Contacting Technical Service	64
8.4	Warnings and Errors List	65
9	Technical Data	70

1 Introduction

1.1 Product Identification

Product:

EM GP2

Automatic Plunge Freezer

Manufacturer:

Leica Mikrosysteme GmbH

Hernalser Hauptstrasse 219

A-1170 Vienna, Austria

www.leica-microsystems.com

1.2 Field of Application and Intended Use

This product, the **EM GP2**, an automatic plunge freezer, is a dedicated cryo sample preparation system for cryo electron microscopy. It was designed and built according to the following European directives:

- Directive 2006/42/EC on Machinery
- Directive 2011/65/EU on Restriction of Hazardous Substances (RoHS)
- Directive 2014/30/EU relating Electromagnetic Compatibility (EMC)

This product delivers a safe and reliable end-user experience, provided it is operated as described in this manual and under the limitations of its technical specifications. Thus, users must read and understand this operation manual before using the product.

Under the conditions stated in this manual, users are responsible for the product's upkeep and shall notify Leica when maintenance and/or repair through authorized Leica service engineers is required. Under no circumstance shall the user modify this product without previous authorization by Leica.

This instrument is for research purposes only; other applications are not approved. This product is neither classified as a Medical Device (European Directive 2007/47/EC), nor intended for In-Vitro Diagnosis (European Directive 1998/79/EC).

This product must not be employed to process any hazardous materials, e.g. radioactive, corrosive, explosive, toxic, or human-pathogenic substances. Leica takes no responsibility for any damage caused by the usage of such substances.

Finally, users must operate this product under the work area conditions outlined in this manual and follow all legally binding safety and environmental regulations.

1.3 Qualification of Operating Personnel

Users, or operating personnel, must be familiar with and follow the recognized rules for safety at work, especially those applying to research laboratories. Users of this product must satisfy the professional requirements for their specific duties.

Users must be familiar with the hazards involving liquefied gases used in cryogenics, the proper usage of personal protective equipment, and the correct measures if accidents occur.

Additionally, users in charge of the product's upkeep (when authorized Leica service engineers are not required) and/or performing additional technical but non-operational tasks (e.g. work area set-up, transportation, decommissioning, and disposal) must be accordingly qualified.

Personnel interacting indirectly with this product (e.g. laboratory personnel working in the proximity) and additional employees without complete knowledge of the safety rules (e.g. administration, caretaking, or cleaning employees) shall be briefed about the safety concerns regarding this product.

Finally, third-parties that are unaware of the recognized safety rules (e.g. external visitors, children) are under no circumstances allowed to operate this product.

1.4 Symbols and Warnings

This manual contains the following signal word panels (according to the standards ISO 3864-2 and ANSI Z535.6) to indicate the severity of hazardous situations.



DANGER indicates a hazard with a high level of risk which, if not avoided, will result in death or serious injury.



WARNING indicates a hazard with a medium level of risk which, if not avoided, could result in death or serious injury.



CAUTION indicates a hazard with a low level of risk which, if not avoided, could result in minor or moderate injury.



NOTICE indicates information considered important but not hazard-related.



SAFETY INSTRUCTIONS indicate specific safety-related instructions or procedures.

In addition to the signal word panels, this manual contains the following pictograms (according to the standard ISO 7010) to indicate the type of hazardous situations.



General Mandatory
Action Sign



Refer to Instruction
Manual



Disconnect Mains Plug
from Electrical Outlet



General Warning Sign



Warning; Freezing Conditions



Warning; Electricity



Warning, Hot Surface



Warning; Flammable Material



Warning, Crushing of Hands

1.5 General Safety Information

Always use this product according to the instructions in this operating manual, and under the conditions described as field of application and intended use (Section 1.2). Additionally, ensure that all relevant operating and non-operating personnel are aware of the possible hazards concerning this product (Section 1.3).

Additionally, do not attempt to bypass the protective measures (Section 2.4), and follow the instructions concerning transportation and storage (Section 3), work area conditions (Section 4), and decommissioning and disposal (Section 7).

The EM GP2 was tested in accordance to IEC/EN 61010-1 and EN 61326-1. It complies with the European Machinery (2006/42/EC), EMC (2014/30/EU) and RoHS (2011/65/EU) directives (Sections 1.2 and 10).


Nevertheless, Leica can only guarantee the device's electrical safety provided that it is installed under the required work area conditions (Section 4.1) and that only authorized Leica service engineers perform all electrical maintenance and repair tasks. Thus, user access to the product's electrical components by removing or bypassing the protective measures is strictly forbidden.



1.6 Safety Measures regarding Nitrogen

1.6.1 Physical Properties

Liquid nitrogen (LN₂) is extremely cold and boils at -196°C. Nitrogen gas (GN₂) escapes at very low temperatures from boiling LN₂. Always consider that both LN₂ and GN₂ as well as cooled elements (e.g. pipes, valves, hoses, containers or stoppers) can cause severe frostbite and burns to the skin and eyes.




When LN₂ evaporates, it expands at a ratio of 1:700. Accordingly, 1 litre LN₂ produces almost 0.7 m³ of GN₂. Therefore, ensure that the room is always well ventilated when large quantities of nitrogen are evaporated (e.g. when transferring LN₂). To dispose of residual LN₂, dump it into an outdoor pit or container filled with gravel, where it will evaporate rapidly and safely.

GN₂ by itself is non-toxic, odourless, tasteless, and is present in air, which – under normal conditions – is composed of 78% nitrogen and 21% oxygen. Nevertheless, additional GN₂ displaces oxygen and can cause immediate fainting and deep unconsciousness without any previous symptoms, such as dizziness and numbness. When the oxygen percentage drops below 18%, death through asphyxiation may occur.

If a person becomes dizzy or loses consciousness while working with LN₂, move him/her to a well-ventilated area immediately. If breathing stops, call an ambulance and apply artificial respiration.



To avoid the above-mentioned hazards, ensure that the laboratory room is always well ventilated. In case of power outage, provide additional sources of ventilation (e.g. opening doors and windows). Additionally, it is recommended to use an oxygen sensor to constantly monitor oxygen levels.




1.6.2 Safe Storage

Prepare storage resources to handle the amount of LN₂ required by this product (plus additional refilling needs) in a safe and appropriate way.

Avoid using glass Dewar vessels, especially those with capacity larger than 2 litres without a complete metal enclosure. Glass Dewars often burst for no apparent reason or due to unintentional mishandling (e.g. contact with metal instruments).


Only use metal Dewars specifically designed for LN₂ storage. Make sure that your Dewar vessel is filled exclusively with LN₂. If necessary, attach a note on it containing “ONLY LIQUID NITROGEN” (or similar) in a visible location.

Check their evaporation rate every three months and compare it with the one given by the manufacturer. The evaporation rate of a metal Dewar in good conditions should be well below 1 litre of LN₂ per day. Defective Dewar vessels with higher evaporation rates present a safety risk and should be removed or subjected to maintenance. Regardless of their state, never leave LN₂ containers in closed storage rooms.



Metal Dewars should only be closed with the provided stoppers, which prevent GN₂ pressure build-up inside the Dewar by providing an adequately big gap for it to escape at regular evaporation rates. Make sure the stopper sits loosely on top of the Dewar neck and that it is not stuck by ice deposits. Do not subject the Dewar to sudden and violent movements, such as shaking, during transportation.

Do not leave the metal Dewars with LN₂ open, as it might mix with oxygen in the atmosphere – LN₂'s boiling point (-196°C) is lower than liquid oxygen's (-183°C). Check the colour of LN₂: a faintly bluish colour indicates a high percentage of liquid oxygen. Liquid oxygen presents additional hazards, as it can cause vigorous burning.



1.6.3 Equipment and Handling

When working with LN_2 , use adequate Personal Protective Equipment (PPE). Avoid closed protective glasses (a), open boots (c), open shoes (e), or protective gloves (g) out of which LN_2 cannot escape easily in case it enters. LN_2 entering the aforementioned equipment does not evaporate quickly and causes serious burns before there is time to remove it.

Instead, always use protective glasses with side protection and openings at the top and bottom (b). Only use boots if you have loose trousers completely covering the gap around them (d). If wearing slip-on sandals (no walking or court shoes), always use cuff less trousers (f).

Never wear protective gloves when pouring LN_2 or when handling with the Dewar vessel. Simply use an open flannel cloth (h) to protect your hands from the cold. Gloves should only be used to handle dry and cold parts and are unsuitable for LN_2 handling.




Figure 1 – PPE when handling LN_2

Insert objects at room temperature (such as forceps) carefully into LN₂ containers, as unexpected boiling or spurting might occur once the object has cooled down.

In case of skin burns from LN₂ splashes, immediately rinse the skin with plenty of water at hand temperature. In case of serious burns, arrange for a skin specialist to see them at once. In case of LN₂ affecting the eyes, immediately rinse with water at hand temperature and arrange for an eye specialist to see them at once.



SAFETY INSTRUCTIONS REGARDING LIQUID NITROGEN

General Instructions

- Follow laboratory rules and binding regulations
- Complete laboratory safety training
- Communicate every safety concern
- Keep safety instructions always for reference

Physical Properties

- Understand the hazards of LN₂ and GN₂
- Dispose residual LN₂ accordingly
- Ensure rooms with LN₂ are always well ventilated
- Provide additional ventilation in case of power outage
- Install an oxygen sensor for additional safety
- Contact emergency services if in immediate danger

Safe Storage

- Do not work alone
- Avoid using glass Dewar vessels
- Use only metal Dewar vessels designed for LN₂
- Control the evaporation rate of the Dewar regularly
- Close metal Dewars only with the loose fitted stoppers
- Check if liquid oxygen is present (bluish colour)
- Fill your Dewar exclusively with LN₂

PPE and Handling

- Handle LN₂ always with care
- Identify and provide adequate PPE
- Employ protective glasses with opening at the bottom
- Wear boots if trousers cover the gap around them
- Use cuff less trousers if wearing slip-on sandals
- Protect your hands with open flannel cloths
- Insert warm objects into LN₂ carefully
- After skin or eye contact rinse with water and contact a specialist

1.7 Safety Measures regarding Ethane and Propane

1.7.1 Physical Properties

This product can use **ethane (C_2H_6)** and **propane (C_3H_8)** as secondary cryogenics. Respectively, they have their melting points at $-183^{\circ}C$ and $-188^{\circ}C$ and their boiling points at $-89^{\circ}C$ and $-42^{\circ}C$. In contrast to LN_2 , liquid ethane and propane do not evaporate immediately upon contact. Therefore, avoid direct contact with these liquefied gases, as they can cause severe frostbite.

When liquid ethane and propane evaporate, they expand at a ratio of 1:430 and 1:310, respectively. In their gaseous state, ethane and propane are colourless, odourless, and flammable. Exposure can cause headache, nausea, vomiting, dizziness and light-headedness. Similarly to LN_2 , very high levels can cause death by asphyxiation from lack of oxygen.

At room temperature, they can form an explosive mixture when combined with air at volume fractions from 2.4% to 14.8% (for ethane) and 1.7% to 10.8% (for propane). Thus, 2.5 ml of these secondary cryogenics on their liquid state can expand into approximately 1 litre of gas, which can cause an explosive reaction in a completely closed environment with a volume of around 7 litres.

Therefore, ensure that the room is always well ventilated. In case of power outage, provide additional sources of ventilation (e.g. opening doors and windows). Additionally, make sure there are no ignition sources in the laboratory (e.g. sparks or Bunsen burners) which can cause an explosion.



1.7.2 Safe Storage

As gaseous ethane and propane are heavier than air, they sink to the ground and can form an explosive mixture in cellars. Therefore, do not use nor store ethane and propane in cellars. If laboratories are situated on ground floors or above, small amounts of gas will flow out under the door and present no hazard.

Ethane and propane storage (pressure) vessels must always be stored in a suitable safety storage container connected to a ventilation system. A gas detection sensor may be used for additional safety.

1.7.3 Equipment and Handling

This product's secondary cryogen container has a volume of 2.5 ml. When working with ethane and propane, use adequate Personal Protective Equipment (PPE). The same kind of equipment required to handle LN_2 can be used to handle liquid ethane and propane (Section 1.6.3).

	NOTICE
	Equipment Follow same considerations as with LN₂ when using ethane or propane.


Similarly, always handle liquid ethane and propane with care. Avoid direct contact at all cost. In case of contact and burns, immediately rinse the affected areas with water at hand temperature and arrange for a specialist to see them at once.

	 WARNING
	Frostbite Rinse with water and consult a specialist in case of skin burs or eye contact.

SAFETY INSTRUCTIONS REGARDING LIQUID ETHANE AND PROPANE	
General Instructions • Follow laboratory rules and binding regulations • Complete laboratory safety training • Communicate every safety concern • Keep safety instructions always for reference Physical Properties • Understand the hazards of ethane and propane • Ensure rooms with ethane and propane are well ventilated • Provide additional ventilation in case of power outage • Avoid having sparks or flame sources in the room Safe Storage • Do not work alone • Do not use or store ethane and propane in cellars • Keep storage vessels in a suitable safety storage containers • Check the storage containers for leakage on a regular basis PPE and Handling • Handle ethane and propane always with care • Identify and provide adequate PPE • Employ protective glasses with opening at the bottom • Wear boots if trousers cover the gap around them • Use cuff less trousers if wearing slip-on sandals • Protect your hands with open flannel cloths • After skin or eye contact rinse with water and contact a specialist	

2 Product Description

2.1 General Overview



- 1 Environmental chamber with humidity and temperature control
- 2 Humidifier tank
- 3 Drip tray
- 4 Magnetic holder for filter paper
- 5 Forceps with forceps adapter
- 6 Viewing system (optional)
- 7 Moveable ring for contamination protection
- 8 Freezing chamber with secondary cryogen and cryo transfer container
- 9 Secondary cryogen liquefier
- 10 Cryo transfer container
- 11 Foot switch (optional)
- 12 Touch panel

The EM GP2 is a dedicated cryo-preparation system for plunge freezing cryo-fixation of thin samples on EM grids. The instrument was designed to make the preparation of vitreous frozen thin films straightforward and reproducible.

All processes are controlled via the touch panel where actual and set parameters can be viewed and the system can be set up according to the user's needs. Parameters can also be saved to programs and recalled at a later point.

An environmental chamber provides controlled temperature and humidity around the grid prior to freezing. An LN₂ Dewar provides cooling to liquefy ethane or propane ("secondary cryogen"), while heaters maintain the secondary cryogen at a suitable temperature and in an environment free of contamination.

Vitrification of a single grid can be completed in within less than a minute, allowing easy upscaling of specimen preparation.

The grid is picked up with forceps fitted with a specialized adapter and connected to the instrument in a defined orientation. The environmental chamber providing controlled humidity and temperature is subsequently lowered around the grid before a small amount of the sample is applied manually onto the grid.


Most of the sample needs to be blotted away with filter paper, to yield a water layer in the 100 nm thickness range. This can be done either by blotting once from one side at a pre-calibrated position of the filter paper, or at a position determined by a blotting sensor, or via multiple blots from the front and/or back side of the grid.

After blotting, the grid carrying the thinned sample is plunged into a secondary cryogen, typically liquefied ethane or propane. Subsequently, the frozen sample is transferred from the cryogen in a cold and dry GN₂ atmosphere into a cryo-grid box for storage or transfer to the next instrument.



2.2 Dimensions and Weight

The following technical drawing describes the dimensions of the EM GP2.



2.2.1 Product Dimensions

The EM GP2 is delivered in one box with the following dimensions and weight:

- **Width:** 650 mm
- **Depth:** 1030 mm
- **Height:** 700 mm
- **Weight:** 72 kg


The EM GP2 itself has the following dimensions and weight. These dimension values do not include the required wall distance during installation (Section 4.1).

- **Width:** 336 mm
- **Depth:** 445 mm
- **Height:** 883 mm
- **Weight:** ~ 45 kg

2.2.2 Stereomicroscope

The dimensions of the basic unit with an installed stereomicroscope are indicated in the following drawing:

- **Max Width:** 675 mm (Furthermost side position)
- **Max Depth:** 805 mm (Furthermost front position)



2.3 Consumption and Emissions

The EM GP2 has the following approximate consumption and emission rates:

- **Voltage:** 100 – 230 V
- **Power:** 220 W
- **LN₂ Dewar capacity:** 1 l
- **LN₂ for initial filling** 1.8 l
- **LN₂ consumption:** ~ 1 l/h ¹
- **Sec. cryogen capacity:** 2.5 ml

2.4 Protective Measures

To improve safety by reducing the occurrence and severity of residual hazards, the EM GP2 is designed with constructive, technical, and informative protective measures. These measures are expected to function throughout the product's lifetime without monitoring nor inspections.

2.4.1 Constructive Protective Measures

The EM GP2 contains a protective case (housing), separating the user from the electrical and mechanical components that are only accessible to authorized Leica service engineers. Openings for air flow are protected. These protective components must be in place during operation.

2.4.2 Technical Protective Measures

As indicated on Section 2.4.3, the chamber's vertical movement might pose a hand (or fingers) crushing hazard. To lower the occurrence and severity of this risk, the following technical protective measures were implemented:


- The chamber motor's force was reduced to a process-acceptable minimum amount. Additionally, the device offers the possibility to decelerate the chamber when it approaches the Dewar area.
- The chamber's movement can be stopped at any time by pressing the red "STOP" button on the centre of the touch panel.



¹ GN₂ flow 50%, Dewar covered with lid, secondary cryogen temperature -180°C

2.4.3 Informative Protective Measures

This operating manual – with the provided descriptions of residual hazards herein – and the safety signals placed on the product are regarded as informative protective measures and shall be kept as such for further reference. The EM GP2 contains the following pictograms and safety signals.



The installed pictograms and safety signals are explained in the following table:

1) Warning, Crushing of Hands



This pictogram is located on the chamber's frame and indicates a hand crushing hazard caused by the chamber's movement. See Section 2.4.2.

Crushing of Hands: Do not place your hand near the chamber while it's moving!

2) Warning, Freezing Conditions



This pictogram is located on the front of the Dewar and alerts about the hazards regarding liquefied gases. See Sections 1.6 and 1.7.

Freezing Conditions: Follow the safety instructions in Sections 1.6 and 1.7!

3) Warning, Suffocation Hazard




This safety signal is located on the right-hand side of the device and warns about the hazard caused by the evaporation of liquefied gases. This hazard is explained in section 1.6.

Suffocation Hazard: Use in a well-ventilated area, even during power outage!

2.5 Type Label

The type label is located on the rear side of the product, above the mains supply switch and cable connectors. See Section 9 for all technical specifications.



3 Transportation and Storage


3.1 Transportation

To prepare the EM GP2 for transport over large distances, empty the humidifier tank and the drip tray (Section 5.4.3), and remove the secondary cryogen container (Section 5.7). If required, perform a bake-out (Section 5.7) to remove any leftover LN₂ from the Dewar and residual moisture. Afterwards, turn off the device and unplug it from the power supply. Finally, contact service or your sales representative to organize the packing and transportation of the EM GP2.



3.2 Relocation

Relocation, i.e. transport within the same or adjacent buildings without needing to pack the product (i.e. transportation using lifts and/or ramps) shall be performed by qualified personnel (Section 1.3). While having the EM GP2 on a lockable wheeled table (Section 4.1), perform the preparation steps in Section 3.1. Make sure that the product stands firmly on the lockable wheeled table – if going through a ramp, two persons are necessary to avoid tilting.



3.3 Storage

In cases when the device needs to be kept un-operational for a prolonged period, clean the humidifier tank (Section 6.3.1) and store the EM GP2 in a room with a temperature between +5 and +35°C (ideally +21 to +24°C) and a maximum relative humidity of 55% (ideally around 40%). Do not connect the device to the mains power supply and make sure that it stays dry throughout the storage time.

4 Installation

4.1 Work Area Conditions

Install the EM GP2 in a room with a temperature between +15 and +30°C (ideally +21 and +24°C) and relative humidity from max. 55% (ideally around 40%). The room must be ventilated, dry, free of dust and other contaminants, and illuminated. Do not install the EM GP2 in area with strong air draught. Do not operate the EM GP2 in outside environments.

Under consideration of the safety measures in Sections 1.6 and 1.7, install the EM GP2 in a well-ventilated room of adequate volume. Make sure that the ventilation system is running at all times. Additional air exchange mechanisms (e.g. door or window opening) must be available if the ventilation could fail during a power outage. Take special care if installing the EM GP2 in a cellar.



The EM GP2 must be installed firmly on a stable, non-slip, dry, even, and vibration-free surface with minimum dimensions of 500 x 700 mm. A lockable, wheeled table with the aforementioned dimensions, with proper height and space for ergonomic sitting, is highly recommended. Additionally, install the instrument with a wall distance of at least 100 mm on the rear side and 300 mm (for sample application) on the lateral sides. Consider the additional space that the stereomicroscope occupies (Section 2.2). The front side must not protrude from the bench.



Additionally, ensure that the work area offers the required mains supply voltage of 100 - 230 V ($\pm 10\%$) and that it is equipped with the required safety measures and personal protective equipment (Section 1.6). Installing an oxygen sensor in the room is highly recommended.

Make sure there is enough space in the room to handle and store liquid nitrogen and ethan/propane comfortably. In case there are other devices working nearby, assure that they not contaminate nor influence the EM GP2 in any negative form.

4.2 Unpacking


Both, the EM GP2 and the optionally included stereomicroscope, are delivered inside a carton box with the dimensions and weight specified on Section 2.2. After delivery, examine the packed EM GP2 and immediately report any transport damage. Additionally, ensure that the product was delivered on the upright position (according to the package) and that it was kept that way.



When required, allow only qualified personnel (Section 1.3) to transfer the packed product to the work area (Section 4.1). These personnel shall ensure that the product is moved safely. Two persons are required to lift the EM GP2!




During installation, an authorized Leica service engineer will unpack the device on the selected work area (Section 4.1) before installation.



The packing material is designed explicitly for the EM GP2 and shall be retained in case the instrument needs to be transported in the future. The product can sustain damage if it is not transported in the original packing. Store the packing material on a dry environment.

4.3 Assembly

An authorized Leica service engineer will assemble the EM GP2 for the first time.



If required (e.g. after relocation), follow the next steps to ensure that the device is properly assembled:

- Level the instrument by adjusting the four adjustable feet (1)
- Connect the touch panel to the respective plug at the rear of the EM GP2 (2)
- Optionally, connect the foot switch to the round plug (3)
- Attach the power cable and connect to mains power supply (4)



If purchased together, the swivel arm for the viewing system will be already attached to the EM GP2. If the viewing system is purchased at a later date, screw the swivel arm onto the underside of the main unit with the bolt supplied. Afterwards, attach the focus arm and lock it with a hex wrench. Finally, attach the stereomicroscope, lock it in place, and locate the eyepieces.



4.4 Accessories

The main outfit of the EM GP2 consists of the basic instrument with a touch panel and an accessories kit as shown in the next figure. During the installation process, the authorized Leica service engineer will corroborate the complete delivery of these accessories and initial consumables.



- 1 Special forceps with insulation coating
- 2 Foam cover
- 3 Styrofoam box and magnetic rings (2 pcs.)
- 4 Filter paper
- 5 Cryotool
- 6 Forceps with forceps adapter (2 pcs.)
- 7 Cryo-grid box (2 pcs.)
- 8 Secondary cryogen container
- 9 Secondary cryogen container cover
- 10 Secondary cryogen gas liquefier
- 11 Cryotransfer container

The single items of this accessories kit can be ordered separately. Please contact your sales representative for a comprehensive accessories and consumables list and the respective order numbers.

5 Operation

5.1 Switching on and User Interface


Turn on the power with the main switch at the rear of the main unit. After the instrument boots up, the following screen will appear:




After touching the screen, the instrument continues to initialise.



After a successful initialisation, the instrument displays the warning messages to a) replace filter paper, b) refill the humidifier tank and remove water from the drip tray, and c) to cool down the instrument by filling the LN₂ container. These steps are described in detail in Section 5.4.



After confirming these warnings (with “OK”), the Main Screen will be shown:



The Main Screen contains the following panels and functional keys:

- 1 The status panel displaying environment, filter paper, and cryogen parameters, alternating periodically between actual and set values.
- 2 The process panel, where the individual steps of the freezing/plunging workflow can be activated in sequence.
- 3 Functional keys, with which special screens or functions can be called, e.g. rotation of the grid around a vertical axis or chamber illumination on/off.
- 4 Warnings and error messages.
- 5 Name and status of the current program.


5.2 Parameters

The freezing parameters can be accessed either via one of the “pen” () buttons on the Main Screen or via the “Edit” button on the list of programs (Section 5.3.2).

The pre-set parameters must be optimised for each sample and support film. Typical settings for each parameter are explained and given in the respective sections below as starting points for optimisation.


5.2.1 Environment Parameters

The “Environment” screen allows to change the parameters related to the environmental chamber (temperature, relative humidity and window heater) and the cryogen (temperature and GN₂ flow).




5.2.1.1 Chamber Temperature


“Chamber Temperature” displays the actual temperature in the environmental chamber, the current set point, and the value saved in the currently selected program which can differ from the current set point if the program has not been saved after the last change (see Section 5.3.2 on saving programs).



Press “Chamber Temperature – Set” to change the current value on the key pad (between 4.0 and 60.0 °C with 1.0 °C steps) and “OK” to confirm.




The chamber temperature regulation can be switched on and off by pressing the toggle button besides “Set” (grey: switched off; green: switched on).



5.2.1.2 Chamber Rel. Humidity

“Chamber Rel. humidity” displays the actual humidity in the environmental chamber, the current set point, and the value saved in the currently selected program which can differ from the current set point if the program has not been saved yet. Press “Chamber Rel. humidity – Set” to change the current value on the key pad (between 1 and 99 % relative humidity with 1 % steps).


The humidity should be chosen as high as possible, particularly at room temperature and above, to avoid evaporation of the thin water layer left after blotting and subsequent changes in concentration of solutes in the sample. Generally, low humidity to promote evaporation yielding thinner ice is not recommended.



Like for the temperature control the humidity regulation can be switched on or off with the button besides “Set”. When it is switched off, the warnings regarding the humidifier tank filling level (see below) will not be displayed either.

5.2.1.3 Chamber window heater

The power of the window heater can be adjusted with the “+/-” buttons in 5% steps. The function is to prevent misting of the bottom part of the front window. The setting shall be adjusted to the actual environmental parameters (temperature / humidity) in the chamber.



The window heating will only become active at 25 °C and above, and not all moisture might be removed from the window at very high humidity.

5.2.1.4 Cryogen temperature


The temperature of the secondary cryogen is regulated to prevent it from freezing at LN₂ temperature. After pressing “Cryogen Temperature – Set”, key in the desired temperature of the secondary cryogen (the “-” sign is automatically preselected), press OK and the value is applied.



The secondary cryogen temperature is best set to the value where it is coldest but still completely liquid (melting points: -183°C for ethane, -188°C for propane). This point can be adjusted prior to plunging and may vary from one gas supply to another, depending on its quality.

5.2.1.5 GN₂ Flow

The Dewar contains a heating element inside the LN₂ which is responsible for producing a dry N₂ gas atmosphere in the freezing area via constant evaporation of LN₂. It can be adjusted with the +/- buttons in 5% steps. It is recommended to use 100% power, especially in laboratories with high humidity, unless the LN₂ consumption in the Dewar needs to be reduced.




The GN₂ function has an automatic control function for certain phases of the process: When the chamber is in the upper position, the flow determined by the above parameter is produced. When the chamber is in the lower position the GN₂ flow is constantly 10%. During the first phase of the up-movement of the chamber the GN₂ flow is switched to 100% to accommodate for the volume increase within the grey ring which moves up with the chamber.

5.2.2 Load Specimen Parameters

The “Load Specimen” screen controls the vertical 180° rotation of the forceps holding the grid before and after applying the specimen through the ports on the left or right side. This rotation allows a flexible workflow for both left- and right-handed operators.




The forceps with the empty grid are always attached to the instrument in the “home” position. The first checkbox (“rotate grid before specimen application”) rotates to a 180° position before specimen application. The second checkbox (“rotate to home position after specimen application”) rotates after specimen application, but only if the grid was away from the home position.



In the example above, a right-handed user prefers to apply the specimen onto the side of the grid with the support film and wants to blot from the same side: the grid is picked up to have the support film facing left when the forceps are connected to the instrument. The grid is then rotated 180°, the right port is used for specimen application, the grid rotates back 180° and is blotted from the left side.

A delay between application of the specimen/pressing, the “Blot” button, and the actual filter paper blotting can be set with the “Delay time before blotting – Set” button (“pre-blotting time”, 0 s to 600 s, 0.1 s steps). This waiting time can be useful to allow adsorption of the specimen to the support film in a controlled environment, typically 0 to 60 s are used.




5.2.3 Blot Parameters

5.2.3.1 Single Blotting without Sensor


In this mode, the filter paper will advance to the user calibrated position of the grid for blotting once, remain there for a defined time, retract, and then either proceed with the freezing process automatically or wait for the user to initiate freezing.

To select this mode, activate “single blotting” and deactivate “sensor blotting”.



The blot time (actual contact of filter paper with grid) can be set up by pressing “Blot time – Set” and keying in the desired blot time (0 to 200 s). This parameter is highly specimen specific – for 3 to 5 μ l aqueous solution, 0.5 to 3.0 s blotting are usual.

The horizontal position of the grid used for blotting can be defined by pressing “Horizontal blot position – no sensor – Set”. If the environmental chamber is in the raised position, a warning will be displayed that the chamber needs to be lowered.



Press “Ok” and the chamber lowers. A “STOP” button appears in case the movement needs to be halted.


The following screen allows to define separate blot positions for the “home” (H) and the “rotated” position of the forceps. Make sure that the calibrated orientation is the same as the actual blot orientation of the grid.



To move the filter paper to the blot position, press “Blot” for the relevant orientation and use the left/right arrows to move the filter paper to the correct position in small steps (0.2 mm). This is when the filter paper is touching all the grid and preferably the tips of the forceps too. The stereomicroscope can help with determining this position accurately.

To set the blot position for the other orientation of the forceps (relevant only for multiple blotting, see section below), swing the blotting arm back to “Home” and then activate “Blot” for the other orientation. The forceps will rotate automatically. To leave this screen, press “Ok”.

The vertical position of the grid relative to the filter paper while blotting can be set starting from the main blotting menu (page 29) by pressing “Set” for the “vertical blot position”.




Also for this step, the environmental chamber must be lowered.

Press “Blot” to move the filter paper to the blotting position. By pressing up/down arrows, the blotter mechanism can be raised or lowered to the desired position. The recommended position is where the upper edge of the grid and the upper edge of the filter paper are aligned with each other.




It is also possible, however, to blot only half the grid or to place the grid into a more central position on the filter paper. To leave this screen, press “Ok”.

5.2.3.2 Single Blotting with Sensor




This is the most automatic setting, intended to increase reproducibility: around the expected position of the grid, the advancement of the filter paper will slow down and a blotting sensor will constantly check for the point where the filter paper is getting wet and starts blotting the sample. Irrespective of the warping of the filter paper (due to the high humidity in the environmental chamber) or bending of the grid, this point standardises the distance from the filter paper to the grid, provided that the sample volume and spreading of the liquid are kept constant.

From the point the blotting sensor is triggered on, the filter paper moves forward by a user defined distance, to exert a reproducible force on the grid. The blotting sensor requires a volume higher than 3 μl on the grid and might not work with solvents other than water.



Sensor blotting requires two parameters to be set up: the “blotting window” for the sensor and the additional movement after the sensor signal has been triggered. To set up the blotting window, select “Blotting window for sensor blotting – Set” to get to the following screen with the chamber lowered.




Insert forceps with an EM grid, rotate the grid to the position that will be used for blotting (home/180°), go to the “Center” position and adjust the position of the filter paper with the left/right arrow buttons to touch the full area of the grid.

The default size of the blotting window (search distance) does not require adjustment under normal conditions. It should be noted that the sensor scans through the search distance on both sides of the Center position. Therefore, the blotting window is double the search distance. Use “OK” to return to the previous screen.

The “additional move” determines the distance through which the filter paper is moved after the sensor is triggered by the filter paper getting wet. This determines the blotting pressure: using a bigger additional movement, the filter paper will exert more pressure on the grid while blotting.

Select “Additional move – Set” from the main blotting menu to get to the setup screen.




Rotate the grid with the “Rotate” button so that the grid side which shall be blotted faces the entry port to be used for sample application. Apply a sample volume of at least 3 µl onto the grid. If necessary rotate the grid so that the drop faces the filter paper. Test the function of the sensor with the “Liquid contact” button, which will stop the filter paper right after the blotting sensor has been triggered. Then advance to the “Blot end” position and determine the desired pressure of filter paper vs grid with the left/right arrows. This sets the additional move parameter.

Go back to the home position, add more sample to the correct side of the grid, and repeat the test blot with “Blot end” until the desired result can be achieved reproducibly. Press “OK” to leave this screen.


5.2.3.3 Multiple blotting

Multiple blotting can be used to blot the grid more than once on one or both sides. This mode of blotting is preferable for **viscous specimens** that cannot be thinned efficiently by single side blotting. The blotting sensor cannot be used in conjunction with multiple blotting, as the first blot is very likely to bring the volume below the 3 µl limit required by the sensor.

To enable this mode, click “multiple blotting” in the blot parameters screen.



The blot time is set up as described above. The number of blots in the home and the 180° position can be selected by pressing the “Set” button next to the blot numbers.



In the above example, the EM GP2 will blot three times in the home position and three times in the 180° position. The maximum number of blots (home and 180° position together) is 10, as there are no more positions on a fresh filter paper.

In this mode, it is essential to have the horizontal position of the filter paper for blotting calibrated accurately for both, the home position and the 180° position (“Horizontal blot position – no sensor – Set”).

5.2.4 Plunge / Transfer Parameters

This section defines the behaviour of the instrument after blotting.



If “Automatically plunge after blotting” is active, no further user interaction is required to freeze the sample after blotting has finished. The “Plunge” button on the main screen/workflow list is disabled as the instrument plunges automatically after the post-blotting time.

If this function is enabled, an optional delay time between blotting and freezing (“post-blotting time”) can be entered. To set the post-blotting time press the “Set” button next to it and key in the desired time (in seconds). Values of 0 to 200 s are allowed.

The post-blotting time is desired by some operators to allow the water layer even out over the surface of the grid. It requires the humidifier to be set up properly to avoid uncontrolled evaporation of the solvent. For typical applications, this delay is not required.


If “Automatically plunge after blotting” is disabled, the user is required to press either the “Plunge” button on the main screen/workflow list or the foot switch to trigger freezing of the sample at an arbitrary point after blotting.

The freezing position is fixed to 4 mm from the base of the secondary cryogen container. To facilitate the transfer of the grid after plunging or to remove excess liquid secondary cryogen, the grid can be slightly lifted out of the cryogen after freezing (“transfer position”).

Entirely depending on the user’s personal preference, the EM GP2 allows to

- Completely skip the transfer position (“Skip transfer position” enabled)
- Automatically advance to the transfer position (“Skip transfer position” disabled, “Automatically move to transfer position” enabled)
- Manually advance to the transfer position with the “Transfer” button on the Main Screen/workflow list or the foot switch (“Skip transfer position” disabled, “Automatically move to transfer position” disabled)

For the latter two cases, the height of the transfer position as compared to the freezing position can be set (“Transfer position above freezing position – Set”).




Use the “Freeze” and “Transfer” buttons to switch between the individual positions. The transfer position can be changed with the adjust arrows only if “Transfer” is active/pressed. The position can be set in 0.5 mm steps from 0 to 5 mm.

5.3 Programs

Programs are used to save and recall all experiment/sample-specific settings of the EM GP2 as described in the previous section. Up to 20 programs can be saved in the instrument's memory, these settings are preserved when the instrument is switched off.

The program list can be accessed via the "Programs" button on the main screen.




The program list contains the name of the program and the dates of the last use and modification. The currently active program is highlighted in red. To navigate the full list of programs (4 pages), use the arrows on the right side of the list.

5.3.1 Loading Programs

To load a program, go to the program list, select the program you want to use, and return to the main screen.

If changes to the previous program were not saved, a warning will be issued allowing you to save the changes or to discard them.




5.3.2 Editing Programs

Programs can be edited by directly accessing the individual parameters from the main screen as described in section 5.2.

Once parameters have been changed, you can either save the changes to the program ("Save") or you can return to the main screen with unsaved parameters (e.g., parameters that need to be tested or that change frequently such as the blotting time).

Unsaved parameters will be applied in the subsequent freezing runs, and are available when the instrument is restarted, but will be replaced by the stored parameters when the program is explicitly re-loaded.


On the main screen, a warning indicates that there are unsaved parameters.



To access and edit freezing parameters from the program list, the program to be edited must be selected, followed by pressing "Edit".

5.3.3 Copying Programs

To copy a program, go to the program list, select the program you want to copy and press "Copy". Now, select the storage space you want to overwrite with the program or press "Main" to return to the main screen.



5.3.4 Deleting Programs

To remove a program, go to the program list, select the respective program and press “Delete”.




After a confirmation, the program's storage space will be overwritten with the instrument's default settings.

5.4 Preparing the Instrument

5.4.1 Filter Paper

Pick up a fresh filter paper ring (Whatman™ #1), preferably with gloves or forceps, and place the metal ring over it. Holding both together, place them well centred onto the magnetic mount in the environmental chamber.




It should be noted that the automatic blotting depends on the type of filter paper used. The sensor is set to recognize blotting of a drop of at least 3 µl by filter paper of the type Whatman™ #1. Sensor blotting may not work properly if any other filter paper is used.

	NOTICE
	Sensor blotting
The sensor for blotting is adjusted to filter paper Whatman™ #1 and a drop size of > 3 µl	



After setting in the fresh filter paper press the  button on the main screen to reset the counter. The instrument will rotate one ring of filter paper 10 times (blots and test blots for alignment) before the paper needs to be replaced. The number of blots remaining is displayed on the main screen and can be reset by pressing the  button. The counter is automatically reset to zero when the instrument is switched on as it is assumed that a new filter paper will be set in at that point (warning message at start up).


Once all 10 positions on the filter paper are used up, a warning will be displayed.




To maintain reproducibility, it is recommended to allow a fresh filter paper to equilibrate for some time in the humid environmental chamber before being used for blotting.

5.4.2 Humidifier


On start up a warning message reminds to fill the humidifier with distilled water. For the first fill, add 90 ml distilled water via the silicone tubing (valve open) into the humidifier tank mounted on the right-hand side of the environmental chamber: Connect a proper syringe² to the tube with the Luer™ lock and open the valve (valve shown in closed position on top, open on bottom). Then press the syringe piston slowly but steadily to fill in the water.




² Not included in the delivery



When the humidifier tank is filled to its maximum, a warning will be displayed together with a warning beep asking to stop the filling. **If the tank is overfilled, the humidifier may not work because the tube connections to the chamber may get clogged.** After filling, close the valve, remove the syringe, and fix the tube on the clip on the rear side of the humidifier cover.



When the filling of the humidifier tank falls below its minimum level during operation the system will beep and indicate that refilling is needed.



The humidifier stops working at that point and will resume operation only after re-filling. **Re-filling requires only 30 ml distilled water.** Whenever the humidifier tank is refilled, it is recommended to empty the condensate tank (see next section).

The humidifier will be activated only when the chamber temperature is close to the set temperature. At higher chamber temperatures, the mist produced by the humidifier might not be visible.

5.4.3 Emptying the Drip Tray

Condensate building up in the environmental chamber is collected in a drip tray below the environmental chamber. As the liquid level here is not monitored, it should be emptied regularly:

Pull out the drip tray to the right, discard of the condensate, and re-insert it to the original position.




5.4.4 Freezing Chamber and Cool-Down

All components of the freezing chamber must be clean and completely dry before starting to work to avoid contamination of the system or freezing of moving parts.

The quickest way to start work is to cool down all necessary components together.

Position the black secondary cryogen container in its holder. Do not condense the secondary cryogen without the black container, as removal of the secondary cryogen will not be possible before bake-out of the instrument.



Load a square or round cryo-grid box into the grid box cryotransfer container and position the container on its platform at the front right in the freezing chamber.




Now pour liquid nitrogen (LN_2), directly into the EM GP2 Dewar using a Dewar or similar. It will take around 1.8 litres LN_2 to cool the Dewar and fill it to 100%.



Fill until the LN₂ level shows 100%. The boiling of the LN₂ will never subside completely as it is actively being evaporated to generate a humidity free GN₂ atmosphere around the cryogen and the specimen.

During operation of the instrument, an LN₂ level < 100% and below the mesh grid in the freezing chamber is normal. A warning will be displayed when the GN₂ flow cannot function anymore below 25% LN₂ level. At 0% Dewar filling level, the cooling of the secondary cryogen cannot be ensured any longer. Either refill with LN₂ immediately or dispose of the secondary cryogen and bake out the EM GP2. A warning will sound when the temperature of the secondary cryogen rises above -100 °C and starts to evaporate.

When secondary cryogen is in the freezing chamber during refill, do not spill LN₂ over the surface of the secondary cryogen – it might freeze over, leading to pressure build-up and explosion. Use the secondary cryogen container cover provided if necessary.




The freezing chamber parameters can be changed by pressing the  button in the “Cryogen” section of the main screen as described in section 5.2.1.4 and 5.2.1.5.

5.4.5 Filling the Secondary Cryogen


Connect the liquefier to the silicone tubing provided and connect the other end to the pressure reduction valve (e.g. 50 mbar) of the secondary cryogen bottle (ethane or propane). To prevent contamination, keep this tubing as short as possible.

Insert the gas liquefier in the chamber over the secondary cryogen container while cooling down the EM GP2 with LN₂ and wait until the temperature of the cryogen container drops to or close to the set value.



Slowly open the cryogen bottle valve. After a few seconds, a white gas can be seen in the liquefier, subsequently condensation of cryogen forming at the inner lip of the liquefier. Monitor the temperature of the secondary cryogen container and stop condensation if the temperature raises too high. Check the cryogen level continuously looking through the window of the liquefier at a flat angle until the container is filled to the brim.

Close all valves on the cryogen bottle and remove the liquefier. When removing the liquefier, avoid contact with any residual liquid ethane/propane that still might be in the system.




5.5 Freezing a Grid

For cryo specimen preparation and sample loading into the microscope, only work with clean and dry (Leica EM CTD) tools to avoid contamination of the specimen, and only approach the frozen specimen with pre-cooled tools (no boiling in LN₂).

Before starting to freeze samples, check if the environmental chamber has equilibrated at the set humidity and temperature and make sure that the secondary cryogen has reached the set temperature after condensation.

If available, the foot switch can be used instead of the workflow buttons on the touch panel to proceed to the next step in the workflow.




5.5.1 Attaching the Forceps/Grid

Perforated or continuous carbon films on EM grids are routinely used as support films for cryo-TEM of biological specimens. These films can either be home-made or purchased and are typically made hydrophilic by exposing them to a glow discharge just before freezing.

The environmental chamber must be lifted to the “LOAD FORCEPS” position to attach the forceps with the grid. If necessary, press “LOAD FORCEPS” in the workflow list to bring the instrument into this state.


The snap-lock forceps have a quick release mechanism for rapid loading and easy unloading. The forceps fit onto the EM GP2 in only one direction to maintain the orientation of the grid (the side with the holey support film). The position during mounting is the “home” position for the forceps (“H” on adapter facing front) and not the 180° vertically rotated position.

Pick up a (glow discharged) grid in the right orientation corresponding to the settings for the specimen loading (section 5.2.2). Lock the forceps with pushing the black locking sleeve to the first groove, and slip the forceps into position on the gantry of the EM GP2.



5.5.2 Lower Chamber

Next, lower the environmental chamber to enclose grid and forceps by pressing “LOAD SPECIMEN” on the main screen.



The chamber will lower and if a 180° rotation for loading has been set up, the forceps will rotate by 180°.



During movement of the environmental chamber, a “STOP” button will appear on the touch panel. Pressing this button will freeze the movement. To continue, the workflow must be reset via the “LOAD FORCEPS” button.



When the environmental chamber is in the lower position, the shutter closing the feedthrough to the cryogen will wiggle in regular intervals to prevent freezing to the bottom of the chamber.

5.5.3 Adjust Parameters


Adjust any freezing parameters, e.g. blotting time, as required by your experiment (section 5.2) or load the required program (section 5.3.1).

For a first run with a new specimen, prepare a sequence of different blot times to achieve the desired film thickness, e.g. 0.7, 1.0, 1.5, and 2.2 seconds.

5.5.4 Applying the Specimen

To dispense the sample to the correct side of the grid the user can pipette fluid through one of the side ports of the environmental chamber. Either port, located to the left and right, can be used, depending on the user preference. It is important that the forceps orientation (home or 180°) is set correctly (section 5.2.2) for the port through which the sample shall be applied.

Open the side port of the environmental chamber and dispense a 3 - 5 μ l fluid droplet onto the grid. Close the side port immediately to maintain a stable atmosphere.



For applying the specimen, a type of microliter pipette must be chosen that can easily reach the grid without getting caught at the ports.


5.5.5 Blotting

By pressing the “BLOT” button on the main screen (or operating the foot switch), the blotting sequence will be started. Depending on program settings, the forceps will possibly rotate to the “home” position.



If a pre-blot is programmed, the screen will show a progress bar as above. To stop the process during the pre-blot time or the blot itself, press “BLOT STOP”. Once the problem that forced the process to be interrupted has been eliminated, the workflow can be resumed by pressing “BLOT” again.

A timer for the blot time(s) is also shown.



If a post-blot delay has been programmed, a timer will be displayed as well.

5.5.6 Freezing

If “automatically plunge after blotting” is selected in the parameters (section 5.2.4, page 34), the EM GP2 will automatically plunge the grid into the secondary cryogen after blotting (and the elapse of the optional post-blot delay).

If this option is not enabled, the user must press the “PLUNGE” button in the workflow list of the main screen or press the foot switch to trigger plunging.



Directly after plunging, the environmental chamber will rise to allow access to the frozen grid. The grid will remain either in the freezing position (lowest position in the secondary cryogen) or be automatically lifted to the “transfer” position a few mm higher if “automatically move to transfer position” is enabled (section 5.2.4).

5.5.7 Transfer to Cryo Grid Box

After plunging, the forceps holding the sample can be lifted to a slightly higher position by pressing “TRANSFER” (if not done automatically).

Remove the lid of the grid box cryo transfer container, leave it inside the Dewar to avoid contamination. Set the grid box to an open position.


Remove the forceps from the gantry tilting it out forwards, always keeping an eye on the frozen grid not to touch the wall of the secondary cryogen container.



With one quick movement, move the grid from the secondary cryogen to the LN₂ in the grid box cryo transfer container and insert it into the assigned slot of the grid box.

Before closing the grid box cryo transfer container, make sure the level of cryogen in the container is high enough to cover the grid box and to transfer grids. If necessary, top up.


When refilling the LN₂ in the grid box cryo transfer container, do not spill LN₂ over the surface of the secondary cryogen – it might freeze over, leading to pressure build-up and explosion. Use the secondary cryogen container cover provided if necessary.



To commence another run, press “Load Forceps” in the workflow and the environmental chamber will rise for the next run.

5.6 Transferring a Cryo-Grid Box

After completing a cryo-grid box, set the grid box to a closed position and fasten it with a pre-cooled screwdriver. Cover the grid box cryo transfer container with its lid, attach the pre-cooled M4 cryotool and lift the container out of the freezing chamber into an additional Dewar filled with LN₂ waiting next to the instrument.



While the container is held with the M4 cryotool below the LN₂ surface in this Dewar, the grid box can be removed from the container with insulated cryo forceps for further storage without exposing it to the air.

To freeze more samples, load the next dry and open cryo-grid box into the grid box cryo transfer container while still inside the additional LN₂ Dewar. Allow the grid box to cool down until boiling subsides, and refill the cryo transfer container completely with LN₂. Move the grid box cryo transfer container back to the EM GP2 Dewar and resume freezing.

5.7 Shutting Down and Bake Out

Upon completion of freezing, transfer the secondary cryogen container using the insulated cryo forceps and appropriate personal protection equipment into the small Styrofoam container provided. Take the container to an explosion proof fume hood and allow the secondary cryogen to evaporate in a safe place.

Remove any remaining water from the humidifier tank by aspirating it with the syringe via the silicone tube and/or lowering the silicone tube into a beaker.

Dispose of the used blotting paper and leave the door of the environmental chamber open.


On completion of the cryo work, the EM GP2 must be baked out to drive off any LN₂ and moisture which may cause problems on the next use by freezing up moving parts or contaminating the specimen.

Go to the “Settings” screen and press “Bake out”. After confirming a warning that the secondary cryogen container has been removed and the viewing systems has been moved to the left side, you can set the bake out time (typically 1h) and start the process by pressing “Start”.



If the Dewar still contains LN_2 , the bake out timer will not start immediately, but only when the Dewar is completely empty. Therefore, the total bake out time will depend on the LN_2 level and the temperature of the environment.

During the bake-out cycle, surfaces in the environmental chamber and the Dewar can become very hot. Touching these surfaces might result in burns.




The bake out cycle can be stopped at any time by pressing “Stop”.

After the bake out is complete, the environmental chamber will move automatically to the bottom position. The instrument will not power down automatically. Before switching off the system with the main switch at the back, wait until the chamber arrived at the lowest position.

5.8 Settings

This screen allows to set a number of experiment-independent parameters on the instrument and software updates. Press “Main” to go back to the main screen.

5.8.1 Tab “Process”



“Ext. remote alarm on” enables/disables a remote alarm signal via a jack plug on the rear of the main instrument. Upon critical alarms (e.g., risk of evaporation of the secondary cryogen), the potential free contact closes.

To have the LED illumination of the Dewar and the environmental chamber automatically switch on during start up, activate the check box “light on at start-up”. Activate “switch light automatically off” to switch off the LEDs after a user defined time. This time can be set by pressing the “Set” button.

The light can be switched on or off at any time with the “light” button on the bottom of the main screen.

The humidifier is disabled by default during sensor blotting to avoid interference

with the measurement. By unchecking the checkbox  Humidifier off during sensor blotting it can be activated during this phase, as well.

By default, the chamber fan is disabled during plunging to prevent humidity from the environmental chamber being blown into the secondary cryogen. This

behaviour can be changed by un-checking  Chamber fan off during plunge

To minimise the disturbance of the dry gaseous N₂ atmosphere in the specimen preparation area on top of the Dewar, the movement speed of the environmental chamber just above the Dewar can be reduced to approx. 1/3. The check boxes in the lower right part of the “Process” tab allow to enable/disable this slow-down for the down and the up movements separately.

5.8.2 Tab “System”



“Volume” sets the volume of the audible beep.

To change the system time select the button “Change” besides the time/date display.

To re-calibrate the touch screen, select button “Calibrate”. It is recommended to do this procedure only when there is a noticeable deviation between the buttons and the touch-sensitive areas on the screen. A stylus is required for accurate touch screen calibration. At the end of the calibration procedure the system asks for pressing “ESC”. With no key pad available on the touch panel, tap anywhere on the screen to confirm the calibration instead.

When the Screenshot function is enabled, it is possible to record a screen shot to a USB stick connected to the USB port. Just type (preferably with a stylus) in vicinity of the top left corner till you hear a confirming signal.

The “System Information Export” function is enabled only when a USB stick is connected. Type on it to export files containing information about errors and other system status notes. These files are intended to be used by service technicians for failure analysis. Please export them and send them to your service technician if asked to do so.

The “Update” and “Service” buttons will be explained in the next section.

6 Maintenance and Repairs

6.1 Service Menu

The EM GP2 has a Service button under the Settings menu (Section 5.8.2) which allows entering the Service menu. Access to this menu is password-protected and restricted to authorized Leica service engineers. Thus, user access to the Service menu is prohibited and considered as not intended usage (Section 1.2).

Additionally, alterations in the Service menu by non-authorized personnel present additional hazard risks (Section 1.5) with significantly higher occurrence and severity than those possible during normal device operation, as some essential protective measures might be deactivated or bypassed.




6.2 Software Updates

The EM GP2 might be subjected to software updates. These updates shall only be provided by your sales representative or authorized Leica service engineers. Do not alter the update files, as it might cause the system to stop functioning!

6.2.1 Preparing for the Update

To update the software, use a UBS-Stick with a storage capacity of at least 4 GB, and perform the following steps:

1. The upgrade is delivered inside a ZIP file (e.g. "EMGP_UI_20171016.zip")
2. Copy this file into the root directory of the USB-Stick
3. Extract the contents of the ZIP file into the root directory of the UBS-Stick (i.e. on a Windows OS right click the ZIP file and choose "Extract All", and actively select the root directory of the USB-Stick e.g. "F:\")
4. Check the folder structure after unzipping. The files should be located at "F:\Leica\GP2update" (assuming the USB-Stick has the drive letter "F:\")



6.2.2 Performing the Update


To perform the software update, press the “Settings” button on the main screen.



Afterwards, press the “Update” button.



Connect the US-Stick with the right folder structure on the rear side of the panel.



The software components needing a software update are shown in bold, with the current and new software version numbers displayed in the third and fourth columns, respectively.

Software	Part ID	Current Version	New Version	Start updates
Operating System				Yes
User Interface	6564 31-912	01.03.08	01.03.09	Yes
Controller CH	6564 31-901	01.03.07	01.03.08	Yes
Controller DW	6564 31-902	01.03.07	01.03.08	Yes

Choose Controller and click "Update".

0.0%

Back

If there is an available update, always update "User Interface" first!



NOTICE

Update
Always update "User Interface" first!

Click on the button "Yes" on the fifth column to begin the update process. The progress bar will update accordingly.

Software	Part ID	Current Version	New Version	Start updates
Operating System				
 User Interface	6564 31-912	01.03.08	01.03.09	
Controller CH	6564 31-901	01.03.07	01.03.08	
Controller DW	6564 31-902	01.03.07	01.03.08	

Copying UI files...


10.0%

Back

After updating the "User Interface", the system will restart automatically. After restarting, enter the Update screen and – if necessary – continue with "Controller CH" and "Controller DW". Restart the system once more after finishing updating.

6.3 Cleaning and Upkeep

Keep the EM GP2 clean by ensuring the chamber's interior is free of dust and other contaminants. To reduce contamination, it is recommended to use gloves when handling in the interior of the chamber! Before cleaning, switch off and unplug the EM GP2 from the mains power supply.



To clean inside the chamber, simply wipe with a wet cloth. Ethanol may also be used. The glass door of the chamber can also be cleaned using commercially available glass cleaners. Additionally, remove any accumulated dust from the exhaust ventilators and protective nets.

6.3.1 Cleaning the Humidifier Tank


The water tank of the humidifier should be flushed out occasionally with 70% ethanol. Simply fill the chamber with 90ml of 70% ethanol, leave it for 15 minutes, drain, and rinse with distilled water.

If the EM GP2 is left unused for longer periods of time, detach and empty the humidifier tank as described below and store it in a dry place. If detaching the tank is not a viable option, add 10ml 70% ethanol to the tank and leave it until the next usage, when it should be drained and rinsed with distilled water.

To disassemble the water tank for cleaning, remove the cover on the right side of the chamber.



Hold the humidifier tank, loosen the securing screw holding the tank, and detach it downwards.



The system can then be cleaned with 70% ethanol and reassembled in the reverse order shown above. When reconnecting the water tank, make sure the two electrical connections fit into their sockets.

7 Disposal

7.1 Decommissioning

To prepare the EM GP2 for decommissioning, empty the humidifier tank (Section 6.3.1) and drip tray (Section 5.4.3), and remove any leftover cryogen liquid from the LN₂ Dewar and secondary cryogen containers. If required, perform a bake-out (Section 5.7) until the device is completely free of moisture. Afterwards, turn off the device and unplug it from the power supply.

Remove or disassemble all accessories and consumables (Section 4.4) and store them into an appropriate storage container. In case of a temporary decommission, follow the instructions for relocation and storage in Sections 3.2 and 3.3.

7.2 Disposal



As displayed by the type label with the crossed-out wheeled bin symbol, the disposal of the EM GP2 inside the European Union (EU) must be done according to the European Directive 2012/19/EU on Waste Electrical and Electronic Equipment (WEEE). This device falls under the category “Large Equipment” according to Annex III of the directive. Users outside the EU must follow their corresponding regulations for appropriate WEEE disposal.

Disposal of the device must be done by a waste recycling contractor authorized to collect and dispose WEEE.

Before disposing, ensure there are no radioactive, highly corrosive, explosive, toxic, or human pathogenic substances inside the device. Failure to do so might endanger the WEEE recycling company’s personnel. In case these substances cannot be reliably removed, contact your corresponding authorities for further information regarding special waste disposal.



8 Troubleshooting

8.1 About this Manual

This operating manual is intended to provide essential information about the proper operation, maintenance, and troubleshooting of this product throughout its service life. Initial and subsequent users must be read and understand this manual before using the EM GP2.

Do not hesitate to communicate any questions, comments, and feedback regarding the content and quality of this manual to your sales representative. Inform your sales representative if this manual is lost or damaged. Occasionally, be sure to contact her/him to inquiry about new or updated versions of this manual. In some cases, the instruction manual might be available electronically for download.



8.2 Warranty

The EM GP2 is covered by a warranty according to the sales conditions. If functional errors should occur or if the components of the system sustain damage that is subject to warranty coverage during the warranty period, the manufacturer will repair or replace the faulty components following examination thereof.

The manufacturer's warranty covers the system in its original configuration. Only original replacements parts may be used. The manufacturer accepts no liability for damage caused by usage of other replacement parts.

The manufacturer will not accept liability for damage caused by misuse of the system or its use for purposes other than the intended use, nor for damage caused by work on the system that is not described in this manual.

8.3 Contacting Technical Service

During operation, you might encounter the warnings and errors listed in Section 8.4. These software messages are intended to guide the user in situations where operational errors are likely and to indicate component malfunction. In some cases, they can also be triggered by unintended user interaction. Refer to the list for further information regarding their causes and possible solutions. Contact your sales representative if you encounter any situation which is not described in this manual. Contact Technical Service if the functioning of the machine (or the safety) gets compromised due to component failure.

8.4 Warnings and Errors List

The following list contains – sorted by the reference code on the first column– all possible warning and error messages displayed by the EM GP2, whereby

- **Warnings** (coded WXXXX) refer to situations needing certain actions by the user to ensure and/or maintain correct system operation, and
- **Errors** (coded EXXXX) refer to wrong and/or unexpected system conditions that require direct interaction from the user (or service engineer) to be lifted.

Besides stating the type of message, the second column displays the possible interactive choices available to the respective warning or error, whereby

- **<OK>** denotes acknowledgement and confirmation of the message
- **<Cancel>** offers the possibility to interrupt the current action
- **<Yes>/<No>** allow accepting or dismissing a proposal
- **<Autohide>** requires no decision (the message disappears by itself)

The third column displays the warning or error's description (sometimes including system status or short advice) as displayed by the software upon appearance. The fourth column provides additional information leading to the probable causes of the warning or error message.

EM GP2 – Errors and Warnings			
Code	Type / Decision	Message	Further Information
W0001	Warning <OK>	Please change the filter paper at start up.	Shown at start up; the filter paper usage counter is set to 10. Insert a new filter paper.
W0002	Warning <OK>	Humidifier start up information. First fill with 90ml water - refill with 30ml water. Please drain the drip tray.	Shown at start up; fill the water tank of the humidifier with 90 ml water (a smaller amount is needed during operation); empty the drip tray after every refill.
W0003	Warning <OK>	LN ₂ is empty! Please fill LN ₂ container.	Shown at start up; fill the LN ₂ container.
W0005	Warning <OK>	Disconnect the USB-Stick. Wait 2 seconds and reconnect the USB-Stick.	USB-Stick communication failure; reconnect the USB-Stick.
E1002	Error <Autohide>	The water tank has become empty. Please refill with 30ml water. First fill with 90ml water. Please drain the drip tray.	The water tank needs refilling; refill the water tank, empty the drip tray after every refill.
E1003	Error <OK>	Water high level reached! Please stop refilling water and drain the drip tray.	The water tank is full; stop filling the water tank; empty the drip tray after every refill.
W1004	Warning <OK>/<Cancel>	The chamber must be moved down to adjust the blot position.	Shown on the editor screen; <OK> will move the chamber down to adjust the blot position.

EM GP2 – Errors and Warnings			
Code	Type / Decision	Message	Further Information
W1005	Warning <OK>/<Cancel>	The chamber movement was stopped. Press <OK> to continue with the chamber movement.	Shown after pressing “STOP” during chamber movement. <OK> resumes movement.
W1006	Warning <OK>/<Cancel>	The chamber must be moved up to adjust the transfer position.	Shown on the editor screen; <OK> will move the chamber up to adjust the transfer position of the plunger arm.
W1007	Warning <OK>/<Cancel>	The chamber must be moved down to adjust the blot position.	Shown on the editor screen; <OK> will move the chamber down to adjust the blot position.
W1008	Warning <OK>/<Cancel>	The chamber must be moved down to adjust the vertical blot position.	Shown on the editor screen; <OK> will move the chamber down to adjust the vertical blot position.
E1009	Error <Autohide>	The cryogen temperature control is disabled on service main screen.	The cryogen temperature control is disabled in Service menu; contact service engineer.
E1010	Error <OK>	Press “LOAD FORCEPS” to initialize the unit.	Plunging did not finish correctly (timeout); press “LOAD FORCEPS” on the main screen to initialise the system.
E1011	Error <OK>	Cryogen temperature too warm. Refill LN ₂ or remove cryogen container. Press <OK> to stop the warning sound.	The temperature in the freezing chamber is higher than -100°C; the secondary cryogen might evaporate. Refill LN ₂ or remove the secondary cryogen container to a fume hood.
E1012	Error <OK>	Blotting with sensor unsuccessful. Check blotting settings.	The blotting sensor did not recognize grid and filter paper contact due to a) no liquid detected on the grid, b) wrong blot position settings, or c) wrong sensor threshold settings.
W1013	Warning <OK>/<Cancel>	Are you sure you want to delete this program?	Confirmation request to delete a program; <OK> proceeds with deletion, <CANCEL> keeps the program unchanged.
W1014	Warning <OK>	The LN ₂ level is low! The GN ₂ -flow cannot no longer be sustained. To prevent icing in the cryogen, refill LN ₂ . <OK> stops the warning sound.	Shown if the LN ₂ level has dropped beneath 25%. Refill LN ₂ immediately to protect the ice-free environment of the freezing chamber.
W1015	Warning <OK>	At power on the cryogen temperature is below 0°C! Remove the cryogen container and start the bake-out process.	Shown when the system starts after a power-off occurred while the freezing chamber was cold. If the power-off was short, continue using the system; otherwise start bake-out before resuming operation to avoid ice in the freezing chamber.

EM GP2 – Errors and Warnings			
Code	Type / Decision	Message	Further Information
W1016	Warning <Yes>/<No>	Modifications in program number ## - (<i>program name</i>) not saved. Do you want to save the program?	Shown when loading a new program without saving changes in the actual program. Before loading the new program, <Yes> saves the changes in the actual program; <No> discards them.
W1017	Warning <Yes>/<No>	Modifications in program number ## - (<i>program name</i>) not saved. Do you want to save the program?	Shown when loading a new program without saving changes in the actual program. Before loading the new program, <Yes> saves the changes in the actual program; <No> discards them.
E10E1	Error <OK>	Chamber controller communication error. Contact service!	Shown when there's no communication between the user interface and the chamber controller. First, restart the system. If the issue persists, contact a service engineer.
E10E2	Error <OK>	Dewar controller communication error. Contact service!	Shown when there's no communication between the user interface and the Dewar controller. First, restart the system. If the issue persists, contact a service engineer.
W1101	Warning <OK>	The filter paper is used up. Please change the filter paper and reset the counter.	Shown when the blotting counter has reached 10. <OK> quits the warning but does not reset the blotting counter!
W1104	Warning <OK>/<Cancel>	The chamber must be moved down to adjust the blot position.	Shown on the editor screen; <OK> will move the chamber down to enable the adjustment of the blot position.
W1301	Warning <Yes>/<No>	Are you sure you want to terminate the bake-out process?	Confirmation request to terminate bake-out; <Yes> terminates it, <No> proceeds with the bake-out.
W1302	Warning <OK>	Please drain the humidifier water tank. Please drain the drip tray.	Shown before bake-out; reminder to empty the system's water reservoirs.
E1303	Error <OK>	Bake-out terminated. Cryogen temperature sensor defect! Contact service.	The cryogen temperature sensor returned an invalid signal; terminating bake-out. Contact a service engineer.
E1304	Error <OK>	Bake-out terminated. Dewar temperature sensor defect! Contact service.	The Dewar temperature sensor returned an invalid signal; terminating bake-out. Contact a service engineer.
E2001	Error <OK>/<Cancel>	Blotter axis initialization failed. Could not reach end switch 'blotter home'.	The blotter is stalled; remove any obstacles in the chamber and press <OK> to reinitialise.

EM GP2 – Errors and Warnings			
Code	Type / Decision	Message	Further Information
		<OK> start initialization blotter motor.	If the issue persists, contact a service engineer.
E2002	Error <OK>/<Cancel>	Blotter axis is not initialized. <OK> start initialization blotter motor.	The blotter is in an undefined state; remove any obstacles in the chamber and press <OK> to reinitialise. If the issue persists, contact a service engineer.
E2003	Error <OK>/<Cancel>	Blotter axis initialization failed. Could not leave end switch 'blotter home'. <OK> start initialization blotter motor.	The blotter is stalled; remove any obstacles in the chamber and press <OK> to reinitialise. If the issue persists, contact a service engineer.
E2006	Error <OK>/<Cancel>	End switch 'blotter home' not found. <OK> start initialization blotter motor.	The blotter is stalled; remove any obstacles in the chamber and press <OK> to reinitialise.
E3001	Error <OK>/<Cancel>	Shutter axis initialization failed. Could not reach end switch 'shutter open'. <OK> start initialization shutter motor	The shutter motor is blocked; remove any obstacles in the chamber and press <OK> to reinitialise.
E3002	Error <OK>/<Cancel>	Shutter axis is not initialized. <OK> start initialization shutter motor.	The shutter is in an undefined state; remove any obstacles in the chamber and press <OK> to reinitialise.
E3003	Error <OK>/<Cancel>	Shutter axis initialization failed. Could not leave end switch 'shutter open'. < OK > start initialization shutter motor.	The shutter motor is blocked; remove any obstacles in the chamber and press <OK> to reinitialise.
E3006	Error <OK>/<Cancel>	End switch 'shutter closed' not found. < OK > start initialization shutter motor.	The shutter motor is stalled; remove any obstacles in the chamber and press <OK> to reinitialise.
E4000	Error <Autohide>	Chamber motor is in motion.	Shown when a command for chamber movement was issued while the chamber was moving; generally no actions required.
E4001	Error <OK>/<Cancel>	Chamber axis initialization failed. Could not reach upper end switch. <OK> start initialization chamber motor.	The chamber could not reach the top position; remove any obstacles and press <OK> to reinitialise.
E4002	Error <OK>/<Cancel>	Chamber axis is not initialized. <OK> start initialization chamber motor.	The chamber is in an undefined state; remove any obstacles and press <OK> to reinitialise.
E4003	Error <OK>/<Cancel>	Chamber axis initialization failed. Could not leave upper end switch. <OK> start initialization chamber motor.	The chamber is stuck at its top position; remove any obstacles and press <OK> to reinitialise.
E4006	Error <OK>/<Cancel>	End switch DOWN not found. <OK> start initialization chamber motor.	The chamber did not reach the bottom position; remove any obstacles and press <OK> to reinitialise.

EM GP2 – Errors and Warnings			
Code	Type / Decision	Message	Further Information
E5001	Error <OK>/<Cancel>	Rotation axis initialization failed. Could not reach end switch. <OK> start initialization rotation motor.	The forceps rotation motor could not initialise; remove any obstacles and press <OK> to reinitialise.
E5002	Error <OK>/<Cancel>	Rotation axis is not initialized. <OK> start initialization rotation motor.	The forceps rotation motor is in an undefined state; remove any obstacles and press <OK> to reinitialise.
E5003	Error <OK>/<Cancel>	Rotation axis initialization failed. Could not leave end switch. <OK> start initialization rotation motor.	The forceps rotation motor is stuck; remove any obstacles and press <OK> to reinitialise.
E6000	Error <Autohide>	Plunger motor is in motion.	Shown when a command for plunger movement was issued while the chamber was moving; generally no actions required.
E6001	Error <OK>/<Cancel>	Plunger axis initialization failed. Could not reach upper end switch. <OK> start initialization plunger.	The plunger retract did not reach the top position; remove any obstacles and press <OK> to reinitialise.
E6002	Error <OK>/<Cancel>	Plunger axis is not initialized. <OK> start initialization plunger.	The plunger retract is in an undefined state; remove any obstacles and press <OK> to reinitialise.
E6003	Error <OK>/<Cancel>	Plunger axis initialization failed. No upper end switch found. <OK> start initialization plunger.	The plunger retract could not initialise; remove any obstacles and press <OK> to reinitialise.
E6005	Error <OK>/<Cancel>	No forceps found. Check sensor forceps. <OK> start initialization plunger.	No forceps were found; clip in forceps and press <OK> to reinitialise.
E6006	Error <OK>/<Cancel>	End switch DOWN not found. <OK> start initialization plunger.	The plunger retract did not encounter its range end sensor; press <OK> to reinitialise. If the issue persists, contact a service engineer.

9 Technical Data

APPROVAL

Conformity Marking	CE (European Union)
--------------------	---------------------

ELECTRICAL

Mains Voltage	100/115/230 V (+/- 10%), 50–60 Hz
Power Consumption	~ 220 W
Control Voltage	24 V DC

DIMENSIONS AND WEIGHT	WIDTH x DEPTH x HEIGHT	WEIGHT
Basic Unit	336 x 445 x 883 mm	~ 45 kg
Basic Unit with Stereomicroscope	Max. 675 x max. 805 x 883 mm	
Instrument packed (one box)	650 x 700 x 1000 mm	~72 kg

OPERATING CONDITIONS	MIN/MAX	IDEAL
Room Temperature	+15 to +30 °C	+21 to +24 °C
Room Rel. Humidity	Max. 55 %	Around 40 %
Storage Temperature	+5 to +35°C	+21 to +24 °C
Storage Humidity	Max. 55 %	Around 40 %
Room Ventilation	Required at all times during operation	

ENVIRONMENTAL CHAMBER

Temperature Control	+4 to +60 °C
Humidity Control	Work Area Humidity to 99% ¹
Chamber Features	Light (LED), Anti-Fog Heating

¹ (test condition: 22 °C environmental temperature, humidity < 55 %)

DEWAR

LN ₂ Dewar Volume	1 l
Dewar Temperature	-196 °C
Dewar bake-out Temperature	Max. +110°C
LN ₂ Consumption	~ 1 l/h ²

² GN₂ flow 50%, Dewar covered with lid, secondary cryogen temperature -180°C

PREPARATION CHAMBER

Allowed Secondary Cryogen	Ethane / Propane
Secondary Cryogen Container Temperature	0 to -196 °C
Secondary Cryogen Container Volume	2.5 ml

ADDITIONAL DATA

Noise Level	< 60 dB (A)
Tilting Angle	> 10 °
Degree of Protection	IP 20

EC Declaration of Conformity

Leica Mikrosysteme GmbH
Hernalser Hauptstrasse 219
A-1170 Vienna, Austria

Machinery	High pressure freezing device for research purpose
Model	Leica EM GP2
Type	EM GP2
The manufacturer expressly declares that the machinery described above fulfills all the relevant provisions of the following European Union directives:	
EMC directive	2014/30/EU
Machinery directive	2006/42/EG
RoHS directive	2011/65/EU
Standards applied:	
Safety requirements for laboratory equipment	EN 61010-1: 2010
EMC evaluation	EN 61326-1: 2013

Conformity assessment carried out by:
TÜV AUSTRIA SERVICES GMBH, Deutschstraße 10, 1230 Vienna Austria
Notified Body 0408

EMC measurement	Certificate INE-AT/EMV-17/176
Test report for EN 61010-1: 2010	Certificate INE-AT/IT-17/144
Person authorised to compile the technical file: Dr. Paul Wurzinger, Hernalser Hauptstrasse 219, A-1170 Vienna, Austria	

ppa Paul Wurzinger

Vienna, 01/11/2017

Dr. Paul Wurzinger, Prokurist



